



PERRY JOHNSON LABORATORY ACCREDITATION, INC.

Certificate of Accreditation

Perry Johnson Laboratory Accreditation, Inc. has assessed the Laboratory of:

A1 Calibration Laboratory S.A.

TERRUM 25 Condominium, Rio Segundo, Alajuela, Costa Rica

(Hereinafter called the Organization) and hereby declares that Organization is accredited in accordance with the recognized International Standard:

ISO/IEC 17025:2017

This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory quality management system (as outlined by the joint ISO-ILAC-IAF Communiqué dated April 2017):

***Chemical, Dimensional, Electrical, Mass (Weighing Devices, Individual Weight), Mechanical, Thermodynamic, Time & Frequency Calibrations
Thermo Hygrometer, and RPM Digital Measurement Instrument
(As detailed in the supplement)***

Accreditation claims for such testing and/or calibration services shall only be made from addresses referenced within this certificate. This Accreditation is granted subject to the system rules governing the Accreditation referred to above, and the Organization hereby covenants with the Accreditation body's duty to observe and comply with the said rules.

For PJLA:

Tracy Szerszen
President

Initial Accreditation Date:

April 9, 2008

Issue Date:

October 04, 2022

Expiration Date:

December 31, 2024

Perry Johnson Laboratory
Accreditation, Inc. (PJLA)
755 W. Big Beaver, Suite 1325
Troy, Michigan 48084

Accreditation No:

59381

Certificate No:

L22-653

The validity of this certificate is maintained through ongoing assessments based on a continuous accreditation cycle. The validity of this certificate should be confirmed through the PJLA website: www.pjllabs.com



Certificate of Accreditation: Supplement

A1 Calibration Laboratory S.A.

TERRUM 25 Condominium, Rio Segunda, Alajuela, Costa Rica
Contact Name: Felix Hernandez Phone: 506-2440-4010

Accreditation is granted to the facility to perform the following calibrations:

Chemical

| MEASURED INSTRUMENT, QUANTITY OR GAUGE | RANGE OR NOMINAL DEVICE SIZE AS APPROPRIATE | CALIBRATION AND MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (\pm) | CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED |
|--|---|--|--|
| pH Meters ^{FO} | Up to 14 pH | 0.03 pH | Buffers Solutions Method I-30 |

Dimensional

| MEASURED INSTRUMENT, QUANTITY OR GAUGE | RANGE OR NOMINAL DEVICE SIZE AS APPROPRIATE | CALIBRATION AND MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (\pm) | CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED |
|--|---|--|---|
| Caliper ^{FO} | Up to 600 mm | (0.007 L + 6) μ m | Gage Blocks Method I-03 Method I-09 Method I-18 |
| Micrometer ^{FO} (inside) | Up to 305 mm | (0.007 L + 6) μ m | |
| Micrometer ^{FO} (outside) | Up to 400 mm | (0.007 L + 6) μ m | |
| Depth Micrometer ^{FO} | Up to 150 mm | (0.007 L + 6) μ m | |
| Dial Indicator ^{FO} | Up to 50 mm | (0.007 L + 6) μ m | Gage Blocks Method I-02 |
| Steel Rule ^F | Up to 1 000 mm | 0.06 mm | Gage Block Portable microscope OMAX M51B Method I-01 |
| Pin Gages ^F | Up to 25.4 mm | 0.002 mm | Laser Micrometer Method I-04 |
| Measuring Tape ^F | Up to 12 000 mm | 0.8 mm | Standard Steel Rule Method I-01 |
| Protractor ^{FO} | 0° to 90° | 0.1 ° | Angle Gage Blocks Method I-12 Method -35 |



Certificate of Accreditation: Supplement

A1 Calibration Laboratory S.A.

TERRUM 25 Condominium, Rio Segunda, Alajuela, Costa Rica
 Contact Name: Felix Hernandez Phone: 506-2440-4010

Accreditation is granted to the facility to perform the following calibrations:

Electrical

| MEASURED INSTRUMENT, QUANTITY OR GAUGE | RANGE OR NOMINAL DEVICE SIZE AS APPROPRIATE | CALIBRATION AND MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (\pm) | CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED |
|--|---|--|---|
| Hypot Tester ^{FO} | 100 V to 9 000 V | 280 μ V/V + 15 mV | Vitrek 4700 Method I-45 |
| pH Meters ^{FO} | -2 000 mV to 2 000 mV | 0.3 mV | Fluke 753 Method I-11 |
| | 15 °C to 40 °C | 0.5 °C | Fluke 753/K Thermocouple |
| Temperature Calibration, Indication and Control Equipment used with Thermocouple Type T ^{FO} | 0 °C to 400 °C | 0.7 °C | Process Calibrator 753 8.5 DMM 8104 Method I-14 Method I-20 Method I-51 |
| Temperature Calibration, Indication and Control Equipment used with Thermocouple Type R ^{FO} | 50 °C to 1 700 °C | 1.2 °C | |
| Temperature Calibration, Indication and Control Equipment used with Thermocouple Type J ^{FO} | -210 °C to 750 °C | 0.5 °C | |
| Temperature Calibration, Indication and Control Equipment used with Thermocouple Type K ^{FO} | -140 °C to 1 340 °C | 0.4 °C | |
| Temperature Calibration Indication & Control Equipment used RTD Type Pt 385, 100 Ω ^{FO} | -200 °C to 800 °C | 0.23 °C | Process Calibrator 753 Method I-14 Method I-20 Method I-51 |
| Temperature Calibration Indication & Control Equipment used RTD Type Pt 3926, 100 Ω ^{FO} | -200 °C to 630 °C | 0.12 °C | |



Certificate of Accreditation: Supplement

A1 Calibration Laboratory S.A.

TERRUM 25 Condominium, Rio Segunda, Alajuela, Costa Rica
 Contact Name: Felix Hernandez Phone: 506-2440-4010

Accreditation is granted to the facility to perform the following calibrations:

Electrical

| MEASURED INSTRUMENT, QUANTITY OR GAUGE | RANGE OR NOMINAL DEVICE SIZE AS APPROPRIATE | CALIBRATION AND MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (\pm) | CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED |
|--|---|---|---|
| Temperature Calibration Indication & Control Equipment used RTD Type Pt 3916, 100 Ω^{FO} | -200 °C to 360 °C | 0.25 °C | Process Calibrator 753 Method I-14 Method I-20 Method I-51 |
| Temperature Calibration Indication & Control Equipment used RTD Type Pt 385, 200 Ω^{FO} | -200 °C to 630 °C | 0.16 °C | |
| Temperature Calibration Indication & Control Equipment used RTD Type Pt 385, 500 Ω^{FO} | -200 °C to 630 °C | 0.12 °C | |
| Temperature Calibration Indication & Control Equipment used RTD Type Pt 385, 1000 Ω^{FO} | -200 °C to 630 °C | 0.23 °C | |
| Temperature Calibration Indication & Control Equipment used RTD Type Pt Ni 672, 120 Ω^{FO} | -200 °C to 260 °C | 0.14 °C | |
| Temperature Calibration Indication & Control Equipment used RTD Type Cu 427, 10 Ω^{FO} | -100 °C to 260 °C | 0.3 °C | |
| Temperature Calibration Indication & Control Equipment used RTD Type Pt 3926, 100 Ω^{FO} | -200 °C to 630 °C | 0.5 °C | |
| Temperature Calibration Indication & Control Equipment used RTD Type Pt 385, 100 Ω^{FO} | -200 °C to 800 °C | 0.8 °C | |
| Temperature Calibration Indication & Control Equipment used RTD Type Pt 385, 200 Ω^{FO} | -200 °C to 630 °C | 0.8 °C | |
| Temperature Calibration Indication & Control Equipment used RTD Type Pt 385, 500 Ω^{FO} | -200 °C to 630 °C | 0.8 °C | |



Certificate of Accreditation: Supplement

A1 Calibration Laboratory S.A.

TERRUM 25 Condominium, Rio Segunda, Alajuela, Costa Rica
 Contact Name: Felix Hernandez Phone: 506-2440-4010

Accreditation is granted to the facility to perform the following calibrations:

Electrical

| MEASURED INSTRUMENT, QUANTITY OR GAUGE | RANGE OR NOMINAL DEVICE SIZE AS APPROPRIATE | CALIBRATION AND MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (\pm) | CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED |
|--|---|--|---|
| Equipment to Output DC Voltage ^{FO} | 0 mV to 100 mV | 5.8 μ V/V + 0.9 μ V | 8.5 DMM 8104 Method I-11 Method I-19 Method I-38 |
| | 100 mV to 1 V | 4.6 μ V/V + 2.5 μ V | |
| | 1 V to 10 V | 4.8 μ V/V + 4.5 μ V | |
| | 10 V to 100 V | 8.5 μ V/V + 72 μ V | |
| | 100 V to 1 000 V | 25 μ V/V + 250 μ V | |
| Equipment to Output AC Voltage at the listed Frequencies ^{FO} | | | |
| 10 Hz to 40 Hz | 0 mV to 100 mV | 86 μ V/V + 99 μ V | |
| 40 Hz to 200 Hz | 0 mV to 100 mV | 86 μ V/V + 45 μ V | |
| 200 Hz to 2 kHz | 0 mV to 100 mV | 86 μ V/V + 38 μ V | |
| 2 kHz to 20 kHz | 0 mV to 100 mV | 86 μ V/V + 53 μ V | |
| 20 kHz to 100 kHz | 0 mV to 100 mV | 86 μ V/V + 160 μ V | |
| Equipment to Output AC Voltage at the listed Frequencies ^{FO} | | | |
| 10 Hz to 40 Hz | 100 mV to 1 V | 92 μ V/V + 50 μ V | |
| 40 Hz to 200 Hz | 100 mV to 1 V | 92 μ V/V + 31 μ V | |
| 200 Hz to 2 kHz | 100 mV to 1 V | 170 μ V/V + 31 μ V | |
| 2 kHz to 20 kHz | 100 mV to 1 V | 350 μ V/V + 31 μ V | |
| 20 kHz to 100 kHz | 100 mV to 1 V | 930 μ V/V + 31 μ V | |
| 100 kHz to 1 MHz | 100 mV to 1 V | 3.5 mV/V + 120 μ V | |
| Equipment to Output AC Voltage at the listed Frequencies ^{FO} | | | |
| 10 Hz to 40 Hz | 1 V to 10 V | 92 μ V/V + 50 μ V | |
| 40 Hz to 200 Hz | 1 V to 10 V | 92 μ V/V + 31 μ V | |
| 200 Hz to 2 kHz | 1 V to 10 V | 170 μ V/V + 31 μ V | |
| 2 kHz to 20 kHz | 1 V to 10 V | 350 μ V/V + 31 μ V | |
| 20 kHz to 100 kHz | 1 V to 10 V | 930 μ V/V + 31 μ V | |
| 100 kHz to 200 kHz | 1 V to 10 V | 3.5 mV/V + 120 μ V | |
| Equipment to Output AC Voltage at the listed Frequencies ^{FO} | | | |
| 10 Hz to 40 Hz | 10 V to 100 V | 450 μ V/V + 2.7 mV | |
| 40 Hz to 200 Hz | 10 V to 100 V | 450 μ V/V + 2.7 mV | |
| 200 Hz to 2 kHz | 10 V to 100 V | 450 μ V/V + 2.7 mV | |
| 2 kHz to 20 kHz | 10 V to 100 V | 560 μ V/V + 2.7 mV | |
| 20 kHz to 50 kHz | 10 V to 100 V | 1.5 mV/V + 2.7 mV | |



Certificate of Accreditation: Supplement

A1 Calibration Laboratory S.A.

TERRUM 25 Condominium, Rio Segunda, Alajuela, Costa Rica
 Contact Name: Felix Hernandez Phone: 506-2440-4010

Accreditation is granted to the facility to perform the following calibrations:

Electrical

| MEASURED INSTRUMENT, QUANTITY OR GAUGE | RANGE OR NOMINAL DEVICE SIZE AS APPROPRIATE | CALIBRATION AND MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (\pm) | CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED |
|--|---|--|---|
| Equipment to Output AC Voltage at the listed Frequencies ^{FO} | | | 8.5 DMM 8104 Method I-11 Method I-19 Method I-38 |
| 10 Hz to 40 Hz | 100 V to 1 000 V | 450 μ V/V + 2.7 mV | |
| 40 Hz to 200 Hz | 100 V to 1 000 V | 450 μ V/V + 2.7 mV | |
| 200 Hz to 2 kHz | 100 V to 1 000 V | 560 μ V/V + 2.7 mV | |
| 2 kHz to 20 kHz | 100 V to 1 000 V | 1.5 mV/V + 2.7 mV | |
| 20 kHz to 50 kHz | 100 V to 1 000 V | 4.7 mV/V + 2.7 mV | |
| Equipment to Output DC Current ^{FO} | | | |
| | 1 nA to 10 nA | 4 μ A/A + 0.2 nA | |
| | 10 nA to 100 nA | 8 μ A/A + 0.8 nA | |
| | 100 nA to 1 μ A | 12 μ A/A + 1.2 nA | |
| | 1 μ A to 10 μ A | 24 μ A/A + 1.2 nA | |
| | 10 μ A to 100 μ A | 24 μ A/A + 1 nA | |
| | 100 μ A to 1 mA | 24 μ A/A + 7.1 nA | |
| | 1 mA to 10 mA | 24 μ A/A + 69 nA | |
| | 10 mA to 100 mA | 41 μ A/A + 680 nA | |
| | 100 mA to 1 A | 130 μ A/A + 13 μ A | |
| | 1 A to 10 A | 130 μ A/A + 26 μ A | |
| | 10 A to 30 A | 130 μ A/A + 80 μ A | |
| Equipment to Output AC Current at the listed Frequencies ^{FO} | | | |
| 10 Hz to 40 Hz | 0 μ A to 100 μ A | 0.47 % of Reading + 35 nA | |
| 40 Hz to 1 kHz | 0 μ A to 100 μ A | 0.18 % of Reading + 35 nA | |
| 1 kHz to 10 kHz | 0 μ A to 100 μ A | 0.07 % of Reading + 35 nA | |
| Equipment to Output AC Current at the listed Frequencies ^{FO} | | | |
| 10 Hz to 40 Hz | 100 μ A to 1 mA | 0.47 % of Reading + 35 nA | |
| 40 Hz to 1 kHz | 100 μ A to 1 mA | 0.18 % of Reading + 35 nA | |
| 1 kHz to 10 kHz | 100 μ A to 1 mA | 0.07 % of Reading + 35 nA | |
| Equipment to Output AC Current at the listed Frequencies ^{FO} | | | |
| 10 Hz to 40 Hz | 1 mA to 10 mA | 0.18 % of Reading + 240 nA | |
| 40 Hz to 1 kHz | 1 mA to 10 mA | 0.07 % of Reading + 240 nA | |
| 1 kHz to 10 kHz | 1 mA to 10 mA | 0.035 % of Reading + 240 nA | |
| Equipment to Output AC Current at the listed Frequencies ^{FO} | | | |
| 10 Hz to 40 Hz | 10 mA to 100 mA | 0.18 % of Reading + 38 μ A | |
| 40 Hz to 1 kHz | 10 mA to 100 mA | 0.07 % of Reading + 43 μ A | |
| 1 kHz to 10 kHz | 10 mA to 100 mA | 0.035 % of Reading + 24 μ A | |



Certificate of Accreditation: Supplement

A1 Calibration Laboratory S.A.

TERRUM 25 Condominium, Rio Segunda, Alajuela, Costa Rica
 Contact Name: Felix Hernandez Phone: 506-2440-4010

Accreditation is granted to the facility to perform the following calibrations:

Electrical

| MEASURED INSTRUMENT, QUANTITY OR GAUGE | RANGE OR NOMINAL DEVICE SIZE AS APPROPRIATE | CALIBRATION AND MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (\pm) | CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED | |
|--|---|--|---|--|
| Equipment to Output AC Current at the listed Frequencies ^{FO} | | | 8.5 DMM 8104 Method I-11 Method I-19 Method I-38 | |
| 10 Hz to 40 Hz | 100 mA to 1 A | 0.47 % of Reading + 240 μ A | | |
| 40 Hz to 1 kHz | 100 mA to 1 A | 0.19 % of Reading + 240 μ A | | |
| 1 kHz to 10 kHz | 100 mA to 1 A | 0.12 % of Reading + 240 μ A | | |
| Equipment to Output AC Current at the listed Frequencies ^{FO} | | | | |
| 10 Hz to 40 Hz | 1 A to 10 A | 0.13 % of Reading + 150 μ A | | |
| 40 Hz to 1 kHz | 1 A to 10 A | 0.1 % of Reading + 150 μ A | | |
| Equipment to Output AC Current at the listed Frequencies ^{FO} | | | | |
| 10 Hz to 40 Hz | 10 A to 30 A | 0.16 % of Reading + 1.5 mA | | |
| 40 Hz to 1 kHz | 10 A to 30 A | 0.11 % of Reading + 1.5 mA | | |
| Equipment to Output Resistance ^{FO} | | | | 8.5 DMM 8104 Method I-37 Method I-32 |
| | 0 Ω to 1 Ω | 22 $\mu\Omega/\Omega$ + 80 $\mu\Omega$ | | |
| | 1 Ω to 10 Ω | 18 $\mu\Omega/\Omega$ + 130 $\mu\Omega$ | | |
| | 10 Ω to 100 Ω | 14 $\mu\Omega/\Omega$ + 1.1 m Ω | | |
| | 100 Ω to 1 k Ω | 12 $\mu\Omega/\Omega$ + 1 m Ω | | |
| | 1 k Ω to 10 k Ω | 12 $\mu\Omega/\Omega$ + 7.4 m Ω | | |
| | 10 k Ω to 100 k Ω | 12 $\mu\Omega/\Omega$ + 320 m Ω | | |
| | 100 k Ω to 1 M Ω | 18 $\mu\Omega/\Omega$ + 7.1 Ω | | |
| | 1 M Ω to 10 M Ω | 59 $\mu\Omega/\Omega$ + 46 Ω | | |
| | 10 M Ω to 100 M Ω | 58 $\mu\Omega/\Omega$ + 5.7 k Ω | | |
| | 100 M Ω to 1 G Ω | 58 $\mu\Omega/\Omega$ + 0.11 M Ω | | |
| | 1 G Ω to 10 G Ω | 58 $\mu\Omega/\Omega$ + 130 k Ω | | |
| | 10 G Ω to 100 G Ω | 58 $\mu\Omega/\Omega$ + 1.1 M Ω | | |
| | 100 G Ω to 1 T Ω | 58 $\mu\Omega/\Omega$ + 11 M Ω | | |
| Equipment to Output Capacitance ^{FO} | | | 8.5 DMM 8104 Method I-11 Method I-19 Method I-38 | |
| | 0 nF to 1 nF | 2 % of Reading + 0.025 nF | | |
| | 1 nF to 10 nF | 1 % of Reading + 0.05 nF | | |
| | 10 nF to 100 nF | 1 % of Reading + 0.5 nF | | |
| | 100 nF to 1 μ F | 1 % of Reading + 5 nF | | |
| | 1 μ F to 10 μ F | 1 % of Reading + 50 nF | | |
| | 10 μ F to 100 μ F | 1 % of Reading + 0.5 μ F | | |
| | 100 μ F to 1 mF | 1 % of Reading + 5 μ F | | |
| | 1 mF to 10 mF | 1 % of Reading + 50 μ F | | |
| | 10 mF to 100 mF | 1 % of Reading + 0.2 mF | | |



Certificate of Accreditation: Supplement

A1 Calibration Laboratory S.A.

TERRUM 25 Condominium, Rio Segunda, Alajuela, Costa Rica
 Contact Name: Felix Hernandez Phone: 506-2440-4010

Accreditation is granted to the facility to perform the following calibrations:

Mechanical

| MEASURED INSTRUMENT, QUANTITY OR GAUGE | RANGE OR NOMINAL DEVICE SIZE AS APPROPRIATE | CALIBRATION AND MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (\pm) | CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED |
|--|---|--|--|
| Pressure Gauge, Differential Pressure Indicators, Sensors and Transducer ^{FO} | 0 kPa to 2.49 kPa | 0.014 kPa | Ashcroft ATE-2/AM2-1 Transducer Method I-07 Method I-06 |
| | 2.49 kPa to 206 kPa | 0.042 kPa | DPI 150 Method I-08 |
| Vacuum Gauges, Vacuum transducers and Sensors ^{FO} | -75.8 kPa to 0 kPa | 0.042 kPa | |
| Aneroid Sphygmomanometer and Sphygmomanometer with Mercury ^{FO} | 0 kPa to 40 kPa | 0.39 kPa | DPI 104 Method I-42 |
| Pressure Gauge, Sensors and Transducer ^{FO} | 206 kPa to 2 068 kPa | 0.18 kPa | UPM Module Method I-07 |
| | 2 068 kPa to 3 447 kPa | 0.41 kPa | |
| Pressure Gauge ^{FO} | 0 kPa to 6 894 kPa | 0.40 kPa | Module 700P30 Method I-07 Module DP 104 Method I-07 |
| | 6 894 kPa to 20 684 kPa | 0.56 kPa | |
| | 20 684 kPa to 34 473 kPa | 4.5 kPa | |
| | 34 473 kPa to 68 948 kPa | 8.0 kPa | |
| Torque Wrench ^{FO} | 0.5 Nm to 5.6 Nm | 0.75 % of Reading | Transducer 2000-400-2 Method I-16 |
| | 3.4 Nm to 45.2 Nm | 0.75 % of Reading | |
| | 9.0 Nm to 113 Nm | 0.75 % of Reading | |
| | 27.1 Nm to 339 Nm | 0.75 % of Reading | |
| | 271.2 Nm to 2 711 Nm | 0.75 % of Reading | Transducer 2000-14-02 Method I-16 |
| Volume Delivery Instruments (Pipettes) ^F | (20 to 100) μ L | 0.8 μ L | MICROBALANCE RADWAG BALANCES & SCALES XA 21.4Y.M.A PLUS ANALITICAL BALANCE OHAUS PA224 Method ISO 8655-6 |
| | (100 to 1 000) μ L | 1.8 μ L | |
| | (1 000 to 5 000) μ L | 8.9 μ L | |
| | (5 000 to 10 000) μ L | 18 μ L | |



Certificate of Accreditation: Supplement

A1 Calibration Laboratory S.A.

TERRUM 25 Condominium, Rio Segunda, Alajuela, Costa Rica
 Contact Name: Felix Hernandez Phone: 506-2440-4010

Accreditation is granted to the facility to perform the following calibrations:

Mass, Force, and Weighing Devices

| MEASURED INSTRUMENT, QUANTITY OR GAUGE | RANGE OR NOMINAL DEVICE SIZE AS APPROPRIATE | CALIBRATION AND MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (\pm) | CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED |
|---|---|--|--|
| Balance ^{FO} | 0 g to 100 g | (0.45 + 0.003Wt) mg | Class F1 Weight Set Method I-13 |
| | 101 g to 300 g | (0.62 + 0.003Wt) mg | |
| | 301 g to 1 000 g | (1.2 + 0.003Wt) mg | |
| | 1 001 g to 2 000 g | (10 + 0.003Wt) mg | |
| | 2 001 g to 10 000 g | (150 + 0.003Wt) mg | |
| | 10 001 g to 30 000 g | (320 + 0.003Wt) mg | |
| | 30 005 g to 60 000 g | (36 + 0.003Wt) g | |
| | 60 005 g to 425 kg | (210 + 0.003Wt) g | |
| Weight (Mass) ^{FO} | 0.5 g | 0.06 mg | OIML F2, M1, M2 CEM Weight Calibration Procedure Method I-27 |
| | 1 g | 0.07 mg | |
| | 2 g | 0.08 mg | |
| | 5 g | 0.1 mg | |
| | 10 g | 0.14 mg | |
| | 20 g | 0.16 mg | |
| | 50 g | 0.24 mg | |
| | 100 g | 0.46 mg | |
| | 200 g | 0.92 mg | |
| | 500 g | 2.4 mg | |
| | 1 000 g | 4.6 mg | |
| | 2 000 g | 9.6 mg | |
| | 5 000 g | 24 mg | |
| | 10 000 g | 46 mg | |
| 20 000 g | 92 mg | | |
| Liquid Volume Measuring Devices to include Graduated Cylinders, Beakers, Burets, Erlenmeyer, Glass Micro Pipettes, Volumetric Balls, Imhoff Cones, Seraphin Test Measures, Gallon "to contain" container or bucket ^F | (Up to 200) mL | 0.003 mL | DIGITAL BALANCE OHAUS PA2202 DIGITAL BALANCE METTLER TOLEDO MS32001LE Gravimetric Method Method CENAM Technical Guide |
| | (200 to 6 000) mL | 0.46 mL | |
| | (6 000 to 25 000) mL | 2.8 mL | |



Certificate of Accreditation: Supplement

A1 Calibration Laboratory S.A.

TERRUM 25 Condominium, Rio Segunda, Alajuela, Costa Rica
 Contact Name: Felix Hernandez Phone: 506-2440-4010

Accreditation is granted to the facility to perform the following calibrations:

Thermodynamic

| MEASURED INSTRUMENT, QUANTITY OR GAUGE | RANGE OR NOMINAL DEVICE SIZE AS APPROPRIATE | CALIBRATION AND MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (\pm) | CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED |
|--|---|--|--|
| RTD Thermometer ^{FO} | -25 °C to 420 °C | 0.08 °C | PTR 5615 / 1521 |
| | 250 °C to 420 °C | 0.1 °C | Method I-14 |
| Temperature Bath ^{FO} | -25 °C to 300 °C | 0.15 °C | PTR 5613 / 1521 Method I-14 |
| Liquid in Glass Thermometer, Digital Thermometer, Bimetallic Thermometer ^{FO} | -25 °C to 300 °C | 0.3 °C | PTR 5615 / 1521 Dry Well 650 S Liquid Bath TE-10D Method I-14 Method I-51 |
| Oven, Freezer, Furnace, Digital Thermometer with thermocouple. ^{FO} | -25 °C to 650 °C | 1.7 °C | Fluke 753 //TC type J and K / Graphtec GL 220/Dry Well 650 S Method I-14 |
| Digital Infrared Thermometer ^{FO} | -25 °C to 35 °C | 1.8 °C | Liquid Bath TE-10D with blackbody target with TC type K Method I-15 |
| | 35 °C to 400 °C | 0.3 °C | |
| Thermo Hygrometer ^{FO} | 15 °C to 30 °C | 0.3 °C | Standard Thermo Hygrometer 635-1 Method I-42 |
| | 20 % RH to 95 % RH | 2.5 % RH | |

Time & Frequency

| MEASURED INSTRUMENT, QUANTITY OR GAUGE | RANGE OR NOMINAL DEVICE SIZE AS APPROPRIATE | CALIBRATION AND MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (\pm) | CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED |
|--|---|--|--|
| Time and Frequency Simulation ^{FO} | 1 μ Hz to 80 MHz | 3×10^{-5} Hz/Hz | Function Generator 4086 Method I-46 Method I-40 Method I-21 |
| | 0 GHz to 2.4 GHz | 60×10^{-6} Hz/Hz | Frequency Counter C3100 |
| RPM Digital Measurement Instrument ^{FO} | 10 rpm to 90 000 rpm | 0.05 % of Reading | Function Generator 4086 Method I-43 |

1. The CMC (Calibration and Measurement Capability) stated for calibrations included on this scope of accreditation represents the smallest measurement uncertainty attainable by the laboratory when performing a more or less routine calibration of a nearly ideal device under nearly ideal conditions. It is typically expressed at a confidence level of 95 % using a coverage factor k (usually equal to 2). The actual measurement uncertainty associated with a specific calibration performed by the laboratory will typically be larger than the CMC for the same calibration since capability and performance of the device being calibrated and the conditions related to the calibration may reasonably be expected to deviate from ideal to some degree.



Certificate of Accreditation: Supplement

A1 Calibration Laboratory S.A.

TERRUM 25 Condominium, Rio Segunda, Alajuela, Costa Rica
Contact Name: Felix Hernandez Phone: 506-2440-4010

Accreditation is granted to the facility to perform the following calibrations:

2. The laboratories range of calibration capability for all disciplines for which they are accredited is the interval from the smallest calibrated standard to the largest calibrated standard used in performing the calibration. The low end of this range must be an attainable value for which the laboratory has or has access to the standard referenced. Verification of an indicated value of zero in the absence of a standard is common practice in the procedure for many calibrations but by its definition it does not constitute calibration of zero capacity.
3. The presence of a superscript F means that the laboratory performs calibration of the indicated parameter at its fixed location. Example: Outside Micrometer^F would mean that the laboratory performs this calibration at its fixed location.
4. The presence of a superscript FO means that the laboratory performs calibration of the indicated parameter both at its fixed location and onsite at customer locations. Example: Outside Micrometer^{FO} would mean that the laboratory performs this calibration at its fixed location and onsite at customer locations.
5. Measurement uncertainties obtained for calibrations performed at customer sites can be expected to be larger than the measurement uncertainties obtained at the laboratories fixed location for similar calibrations. This is due to the effects of transportation of the standards and equipment and upon environmental conditions at the customer site which are typically not controlled as closely as at the laboratories fixed location.